

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021919\
 Data File : VN053868.D
 Acq On : 19 Feb 2019 15:52
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 20 03:46:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	0.515	0.461	10.5	71	0.00
3 P	Chloromethane	0.594	0.487	18.0	71	0.00
4 C	Vinyl Chloride	0.576	0.523	9.2#	78	0.00
5 T	Bromomethane	0.377	0.379	-0.5	89	0.00
6 T	Chloroethane	0.329	0.351	-6.7	94	0.00
7 T	Trichlorofluoromethane	0.781	0.790	-1.2	88	0.00
8 T	Diethyl Ether	0.285	0.253	11.2	78	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.483	0.478	1.0	88	0.00
10 T	Methyl Iodide	0.759	0.674	11.2	78	0.00
11 T	Tert butyl alcohol	0.030	0.026	13.3	79	0.00
12 CM	1,1-Dichloroethene	0.463	0.421	9.1#	79	0.00
13 T	Acrolein	0.016	0.015	6.3	101	0.00
14 T	Allyl chloride	0.737	0.626	15.1	73	0.00
15 T	Acrylonitrile	0.160	0.147	8.1	79	0.00
16 T	Acetone	0.129	0.127	1.6	92	0.00
17 T	Carbon Disulfide	1.442	1.084	24.8	66	0.00
18 T	Methyl Acetate	0.381	0.389	-2.1	93	0.00
19 T	Methyl tert-butyl Ether	1.174	1.137	3.2	82	0.00
20 T	Methylene Chloride	0.540	0.486	10.0	83	0.00
21 T	trans-1,2-Dichloroethene	0.500	0.451	9.8	79	0.00
22 T	Diisopropyl ether	1.366	1.343	1.7	84	0.00
23 T	Vinyl Acetate	0.945	0.854	9.6	75	0.00
24 P	1,1-Dichloroethane	0.911	0.855	6.1	83	0.00
25 T	2-Butanone	0.183	0.166	9.3	79	0.00
26 T	2,2-Dichloropropane	0.696	0.687	1.3	87	0.00
27 T	cis-1,2-Dichloroethene	0.555	0.535	3.6	83	0.00
28 T	Bromochloromethane	0.395	0.324	18.0	72	0.00
29 T	Tetrahydrofuran	0.118	0.103	12.7	74	0.00
30 C	Chloroform	0.903	0.884	2.1#	88	0.00
31 T	Cyclohexane	0.949	0.743	21.7	78	0.00
32 T	1,1,1-Trichloroethane	0.752	0.755	-0.4	88	0.00
33 S	1,2-Dichloroethane-d4	0.502	0.456	9.2	80	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.316	0.305	3.5	81	0.00
36 T	1,1-Dichloropropene	0.451	0.484	-7.3	90	0.00
37 T	Ethyl Acetate	0.252	0.235	6.7	72	0.00
38 T	Carbon Tetrachloride	0.454	0.479	-5.5	89	0.00
39 T	Methylcyclohexane	0.535	0.543	-1.5	82	0.00
40 TM	Benzene	1.390	1.459	-5.0	88	0.00
41 T	Methacrylonitrile	0.141	0.160	-13.5	95	0.02
42 TM	1,2-Dichloroethane	0.413	0.426	-3.1	88	0.00
43 T	Isopropyl Acetate	0.486	0.462	4.9	82	0.00
44 TM	Trichloroethene	0.396	0.392	1.0	86	0.00
45 C	1,2-Dichloropropane	0.356	0.365	-2.5#	85	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021919\
 Data File : VN053868.D
 Acq On : 19 Feb 2019 15:52
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 20 03:46:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.214	0.210	1.9	84	0.00
47 T	Bromodichloromethane	0.446	0.466	-4.5	88	0.00
48 T	Methyl methacrylate	0.232	0.206	11.2	74	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	82	0.00
50 S	Toluene-d8	1.186	1.103	7.0	78	0.00
51 T	4-Methyl-2-Pentanone	0.239	0.232	2.9	78	0.00
52 CM	Toluene	0.825	0.840	-1.8#	84	0.00
53 T	t-1,3-Dichloropropene	0.446	0.460	-3.1	85	0.00
54 T	cis-1,3-Dichloropropene	0.529	0.544	-2.8	85	0.00
55 T	1,1,2-Trichloroethane	0.297	0.307	-3.4	85	0.00
56 T	Ethyl methacrylate	0.352	0.354	-0.6	80	0.00
57 T	1,3-Dichloropropane	0.508	0.505	0.6	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.196	0.154	21.4	61	0.00
59 T	2-Hexanone	0.162	0.151	6.8	76	0.00
60 T	Dibromochloromethane	0.335	0.358	-6.9	88	0.00
61 T	1,2-Dibromoethane	0.294	0.286	2.7	82	0.00
62 S	4-Bromofluorobenzene	0.396	0.380	4.0	80	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	81	0.00
64 T	Tetrachloroethene	0.429	0.453	-5.6	88	0.00
65 PM	Chlorobenzene	1.062	1.105	-4.0	87	0.00
66 T	1,1,1,2-Tetrachloroethane	0.371	0.411	-10.8	90	0.00
67 C	Ethyl Benzene	1.745	1.830	-4.9#	85	0.00
68 T	m/p-Xylenes	0.666	0.720	-8.1	86	0.00
69 T	o-Xylene	0.643	0.697	-8.4	86	0.00
70 T	Styrene	1.043	1.142	-9.5	87	0.00
71 P	Bromoform	0.254	0.282	-11.0	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
73 T	Isopropylbenzene	3.515	3.591	-2.2	89	0.00
74 T	N-amyl acetate	0.845	0.778	7.9	79	0.00
75 P	1,1,2,2-Tetrachloroethane	0.841	0.758	9.9	81	0.00
76 T	1,2,3-Trichloropropane	0.642	0.724	-12.8	100	0.00
77 T	Bromobenzene	0.947	0.936	1.2	87	0.00
78 T	n-propylbenzene	4.063	4.235	-4.2	90	0.00
79 T	2-Chlorotoluene	2.411	2.449	-1.6	89	0.00
80 T	1,3,5-Trimethylbenzene	2.915	3.113	-6.8	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.228	0.215	5.7	80	0.00
82 T	4-Chlorotoluene	2.474	2.503	-1.2	88	0.00
83 T	tert-Butylbenzene	2.576	2.749	-6.7	92	0.00
84 T	1,2,4-Trimethylbenzene	2.970	3.155	-6.2	90	0.00
85 T	sec-Butylbenzene	3.506	3.830	-9.2	94	0.00
86 T	p-Isopropyltoluene	3.066	3.474	-13.3	96	0.00
87 T	1,3-Dichlorobenzene	1.703	1.808	-6.2	94	0.00
88 T	1,4-Dichlorobenzene	1.685	1.749	-3.8	92	0.00
89 T	n-Butylbenzene	2.654	3.060	-15.3	98	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021919\
Data File : VN053868.D
Acq On : 19 Feb 2019 15:52
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Feb 20 03:46:47 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.506	0.615	-21.5	108	0.00
91 T	1,2-Dichlorobenzene	1.630	1.813	-11.2	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.120	0.116	3.3	83	0.00
93 T	1,2,4-Trichlorobenzene	0.935	0.980	-4.8	85	0.00
94 T	Hexachlorobutadiene	0.582	0.723	-24.2	108	0.00
95 T	Naphthalene	1.905	1.720	9.7	69	0.00
96 T	1,2,3-Trichlorobenzene	0.868	0.911	-5.0	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6