

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN021920\
 Data File : VN060144.D
 Acq On : 19 Feb 2020 10:24
 Operator : JC/MD
 Sample : VN0219MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0219MBL01

Quant Time: Feb 20 01:38:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	170610	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	264709	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	228046	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	82162	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	99956	48.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.30%	
35) Dibromofluoromethane	7.56	113	80785	49.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.30%	
50) Toluene-d8	10.08	98	316620	49.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.82%	
62) 4-Bromofluorobenzene	12.40	95	99144	42.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.22%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN021920\
Data File : VN060144.D
Acq On : 19 Feb 2020 10:24
Operator : JC/MD
Sample : VN0219MBL01
Misc : 5.00µL/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0219MBL01

Quant Time: Feb 20 01:38:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

