

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021920\
 Data File : VN060152.D
 Acq On : 19 Feb 2020 13:46
 Operator : JC/MD
 Sample : PB126910ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB126910ZHE#04

Quant Time: Feb 20 03:03:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	181526	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	290303	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	270879	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	121108	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	106426	48.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.36%	
35) Dibromofluoromethane	7.57	113	85036	47.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.30%	
50) Toluene-d8	10.08	98	349640	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
62) 4-Bromofluorobenzene	12.40	95	127225	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
Target Compounds						
20) Methylene Chloride	4.53	84	7342	3.820	ug/l	94
43) Isopropyl Acetate	8.14	43	48949	11.783	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021920\
Data File : VN060152.D
Acq On : 19 Feb 2020 13:46
Operator : JC/MD
Sample : PB126910ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB126910ZHE#04

Quant Time: Feb 20 03:03:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

