

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021920\
 Data File : VN060155.D
 Acq On : 19 Feb 2020 14:55
 Operator : JC/MD
 Sample : L1573-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LOC-2

Quant Time: Feb 20 03:10:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	171335	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	270936	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	255005	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	113849	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	101274	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
35) Dibromofluoromethane	7.57	113	82577	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
50) Toluene-d8	10.08	98	329718	50.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.54%	
62) 4-Bromofluorobenzene	12.40	95	116087	48.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.34%	
Target Compounds						
16) Acetone	3.80	43	6176	6.791	ug/l	98
20) Methylene Chloride	4.53	84	8319	4.585	ug/l	97
43) Isopropyl Acetate	8.14	43	26543	6.846	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021920\
Data File : VN060155.D
Acq On : 19 Feb 2020 14:55
Operator : JC/MD
Sample : L1573-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LOC-2

Quant Time: Feb 20 03:10:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

