


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN021920\
Data File : VN060142.D
Acq On : 19 Feb 2020 9:14
Operator : JC/MD
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Title : SW846 8260
Last Update : Wed Feb 12 14:12:21 2020



AutoFind: Scans 3303, 3304, 3305; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.9	5078	PASS
75	95	30	60	50.8	12342	PASS
95	95	100	100	100.0	24312	PASS
96	95	5	9	6.8	1645	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	86.4	21013	PASS
175	174	5	9	7.5	1579	PASS
176	174	95	101	96.6	20293	PASS
177	176	5	9	7.3	1477	PASS