

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021925\
 Data File : VN085791.D
 Acq On : 19 Feb 2025 15:32
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Feb 20 00:49:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.218	168	237544	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	441696	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.859	117	395623	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	159848	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	182354	59.212	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 118.420%			
35) Dibromofluoromethane	8.165	113	156626	54.191	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 108.380%			
50) Toluene-d8	10.565	98	543597	51.694	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.380%			
62) 4-Bromofluorobenzene	12.847	95	165235	47.692	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 95.380%			
Target Compounds						Qvalue
3) Chloromethane	2.371	50	668	0.206	ug/l #	41
16) Acetone	4.442	43	223	0.229	ug/l #	40
18) Methyl Acetate	5.048	43	750	0.224	ug/l #	47
31) Cyclohexane	8.224	56	4976	1.131	ug/l #	1
36) 1,1-Dichloropropene	8.224	75	19463	4.758	ug/l #	51
38) Carbon Tetrachloride	8.224	117	25756	5.299	ug/l #	18
43) Isopropyl Acetate	8.694	43	2752	Below Cal	#	53
76) 1,2,3-Trichloropropane	12.847	75	83733	23.582	ug/l #	1
81) trans-1,4-Dichloro-2-b...	12.847	75	83733	68.806	ug/l #	12


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021925\
Data File : VN085791.D
Acq On : 19 Feb 2025 15:32
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Feb 20 00:49:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 1
Delta R.T. -0.000 min
Lab File: VN085791.D
Acq: 19 Feb 2025 15:32




Instrument :
MSVOA_N
ClientSampleId :
IBLK


Tgt Ion:168 Resp: 237544
Ion Ratio Lower Upper
168 100
99 61.7 47.9 71.9



#3
Chloromethane
Concen: 0.206 ug/l
RT: 2.371 min Scan# 71
Delta R.T. 0.012 min
Lab File: VN085791.D
Acq: 19 Feb 2025 15:32

Tgt Ion: 50 Resp: 668
Ion Ratio Lower Upper
50 100
52 0.0 26.7 40.1#









#16
Acetone
Concen: 0.229 ug/l
RT: 4.442 min Scan# 419
Delta R.T. 0.024 min
Lab File: VN085791.D
Acq: 19 Feb 2025 15:32




Instrument :
MSVOA_N
ClientSampleId :
IBLK


Tgt Ion: 43 Resp: 223
Ion Ratio Lower Upper
43 100
58 0.0 27.5 41.3#



#18
Methyl Acetate
Concen: 0.224 ug/l
RT: 5.048 min Scan# 526
Delta R.T. 0.029 min
Lab File: VN085791.D
Acq: 19 Feb 2025 15:32


Tgt Ion: 43 Resp: 750
Ion Ratio Lower Upper
43 100
74 0.0 22.2 33.4#








#31
Cyclohexane
Concen: 1.131 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.029 min
Lab File: VN085791.D
Acq: 19 Feb 2025 15:32

Instrument :
MSVOA_N
ClientSampleId :
IBLK







Tgt Ion: 56 Resp: 4976
Ion Ratio Lower Upper
56 100
69 199.2 25.8 38.6#
84 148.7 74.5 111.7#



#33
1,2-Dichloroethane-d4
Concen: 59.212 ug/l
RT: 8.571 min Scan# 1125
Delta R.T. 0.000 min
Lab File: VN085791.D
Acq: 19 Feb 2025 15:32

Tgt Ion: 65 Resp: 182354
Ion Ratio Lower Upper
65 100
67 52.6 0.0 106.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN085791.D

Acq: 19 Feb 2025 15:32

Instrument :

MSVOA_N

ClientSampleId :

IBLK





Tgt Ion:114 Resp: 441696

Ion Ratio Lower Upper

114 100

63 19.6 0.0 39.2

88 15.6 0.0 30.0



#35

Dibromofluoromethane

Concen: 54.191 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN085791.D

Acq: 19 Feb 2025 15:32




Tgt Ion:113 Resp: 156626


Ion Ratio Lower Upper

113 100

111 103.8 82.2 123.4

192 18.5 14.7 22.1





#36

1,1-Dichloropropene

Concen: 4.758 ug/l

RT: 8.224 min Scan# 1066

Delta R.T. -0.141 min

Lab File: VN085791.D

Acq: 19 Feb 2025 15:32

Instrument :

MSVOA_N

ClientSampleId :

IBLK





Tgt Ion: 75 Resp: 19463

Ion Ratio Lower Upper

75 100

110 9.9 17.3 52.0#

77 0.0 24.6 37.0#



#38

Carbon Tetrachloride

Concen: 5.299 ug/l

RT: 8.224 min Scan# 1066

Delta R.T. -0.129 min

Lab File: VN085791.D

Acq: 19 Feb 2025 15:32




Tgt Ion: 117 Resp: 25756


Ion Ratio Lower Upper

117 100

119 5.8 73.8 110.6#

121 0.0 24.3 36.5#









#50
Toluene-d8
Concen: 51.694 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.006 min
Lab File: VN085791.D
Acq: 19 Feb 2025 15:32




Instrument :
MSVOA_N
ClientSampleId :
IBLK


Tgt Ion: 98 Resp: 543597
Ion Ratio Lower Upper
98 100
100 63.8 52.1 78.1



#62
4-Bromofluorobenzene
Concen: 47.692 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.006 min
Lab File: VN085791.D
Acq: 19 Feb 2025 15:32

Tgt Ion: 95 Resp: 165235
Ion Ratio Lower Upper
95 100
174 79.0 0.0 152.4
176 74.4 0.0 146.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.859 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN085791.D
Acq: 19 Feb 2025 15:32




Instrument :
MSVOA_N
ClientSampleId :
IBLK


Tgt Ion:117 Resp: 395623
Ion Ratio Lower Upper
117 100
82 56.9 45.7 68.5
119 31.3 26.2 39.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN085791.D
Acq: 19 Feb 2025 15:32

Tgt Ion:152 Resp: 159848
Ion Ratio Lower Upper
152 100
115 60.9 30.4 91.3
150 157.1 0.0 345.4









#76
1,2,3-Trichloropropane
Concen: 23.582 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.141 min
Lab File: VN085791.D
Acq: 19 Feb 2025 15:32

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Tgt Ion: 75 Resp: 83733
Ion Ratio Lower Upper
75 100
77 1.3 100.0 300.0#



#81
trans-1,4-Dichloro-2-butene
Concen: 68.806 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.118 min
Lab File: VN085791.D
Acq: 19 Feb 2025 15:32

Tgt Ion: 75 Resp: 83733
Ion Ratio Lower Upper
75 100
53 0.0 74.2 111.4#
89 0.0 34.9 52.3#

