

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021925\
 Data File : VN085784.D
 Acq On : 19 Feb 2025 11:49
 Operator : JC\MD
 Sample : VN0219WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0219WBL01

Quant Time: Feb 19 12:05:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	286037	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	514740	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.859	117	457806	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	185142	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	207977	56.08	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.16%
35) Dibromofluoromethane	8.159	113	183991	54.63	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.26%
50) Toluene-d8	10.559	98	636245	51.92	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.84%
62) 4-Bromofluorobenzene	12.847	95	194403	48.15	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.30%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021925\
Data File : VN085784.D
Acq On : 19 Feb 2025 11:49
Operator : JC\MD
Sample : VN0219WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0219WBL01

Quant Time: Feb 19 12:05:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

