

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022020\
 Data File : VN060173.D
 Acq On : 20 Feb 2020 15:33
 Operator : JC/MD
 Sample : L1610-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T1-SECTION-C

Quant Time: Feb 21 03:22:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	207456	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	326994	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	300303	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	134158	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	123608	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	
35) Dibromofluoromethane	7.57	113	96264	47.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.78%	
50) Toluene-d8	10.08	98	392481	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
62) 4-Bromofluorobenzene	12.40	95	141709	48.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.44%	
Target Compounds						
30) Chloroform	7.35	83	64895	16.163	ug/l	98
47) Bromodichloromethane	9.39	83	7129	2.231	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022020\
Data File : VN060173.D
Acq On : 20 Feb 2020 15:33
Operator : JC/MD
Sample : L1610-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T1-SECTION-C

Quant Time: Feb 21 03:22:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

