

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022020\
 Data File : VN060174.D
 Acq On : 20 Feb 2020 16:07
 Operator : JC/MD
 Sample : PB126926TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB126926TB

Quant Time: Feb 21 03:23:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	188938	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	293172	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	276430	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	125910	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	109978	47.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.68%	
35) Dibromofluoromethane	7.57	113	87452	48.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.06%	
50) Toluene-d8	10.08	98	359494	50.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.32%	
62) 4-Bromofluorobenzene	12.40	95	130904	50.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.40%	
Target Compounds						
20) Methylene Chloride	4.53	84	5106	2.552	ug/l	94
43) Isopropyl Acetate	8.14	43	50066	11.934	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022020\
Data File : VN060174.D
Acq On : 20 Feb 2020 16:07
Operator : JC/MD
Sample : PB126926TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB126926TB

Quant Time: Feb 21 03:23:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

