

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022020\
 Data File : VN060175.D
 Acq On : 20 Feb 2020 16:29
 Operator : JC/MD
 Sample : L1578-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WFD-WC-S-SB-010203

Quant Time: Feb 21 03:25:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	197893	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	309630	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	293322	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	132368	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	115920	48.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.28%	
35) Dibromofluoromethane	7.57	113	93394	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	
50) Toluene-d8	10.08	98	378578	50.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.02%	
62) 4-Bromofluorobenzene	12.40	95	136874	49.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.40%	
Target Compounds						
16) Acetone	3.79	43	10504	10.000	ug/l	93
20) Methylene Chloride	4.53	84	10392	4.959	ug/l	95
43) Isopropyl Acetate	8.14	43	22774	5.140	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022020\
Data File : VN060175.D
Acq On : 20 Feb 2020 16:29
Operator : JC/MD
Sample : L1578-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WFD-WC-S-SB-010203

Quant Time: Feb 21 03:25:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

