

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022025\
 Data File : VN085796.D
 Acq On : 20 Feb 2025 11:44
 Operator : JC\MD
 Sample : VN0220WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0220WBL01

Quant Time: Feb 21 00:57:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	259013	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	473649	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.859	117	428474	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	172706	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	201542	60.018	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	120.040%
35) Dibromofluoromethane	8.159	113	174562	56.322	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.640%
50) Toluene-d8	10.559	98	585576	51.930	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.860%
62) 4-Bromofluorobenzene	12.847	95	176350	47.466	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022025\
Data File : VN085796.D
Acq On : 20 Feb 2025 11:44
Operator : JC\MD
Sample : VN0220WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0220WBL01

Quant Time: Feb 21 00:57:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

