

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022119\
 Data File : VN053884.D
 Acq On : 21 Feb 2019 11:36
 Operator : JC/SP
 Sample : K1540-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB

Quant Time: Feb 22 01:06:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 21 03:36:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	652372	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1060692	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1007853	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	374579	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	328403	49.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.52%	
35) Dibromofluoromethane	7.59	113	336789	48.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.76%	
50) Toluene-d8	10.09	98	1299988	51.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.42%	
62) 4-Bromofluorobenzene	12.40	95	424030	51.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.18%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022119\
Data File : VN053884.D
Acq On : 21 Feb 2019 11:36
Operator : JC/SP
Sample : K1540-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB

Quant Time: Feb 22 01:06:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022019W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 21 03:36:07 2019
Response via : Initial Calibration

