

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022119\
 Data File : VN053886.D
 Acq On : 21 Feb 2019 12:29
 Operator : JC/SP
 Sample : K1540-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 M2-T2-OZ(PRE)

Quant Time: Feb 22 01:12:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 21 03:36:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	637779	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1083099	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1108311	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	393598	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	339899	52.15	ug/l	0.00
Spiked Amount						
			Recovery	=	104.30%	
35) Dibromofluoromethane	7.59	113	345285	48.58	ug/l	0.00
Spiked Amount						
			Recovery	=	97.16%	
50) Toluene-d8	10.09	98	1307311	50.92	ug/l	0.00
Spiked Amount						
			Recovery	=	101.84%	
62) 4-Bromofluorobenzene	12.40	95	476582	56.23	ug/l	0.00
Spiked Amount						
			Recovery	=	112.46%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	11985	6.686	ug/l	95
19) Methyl tert-butyl Ether	5.05	73	14039	0.935	ug/l	100
25) 2-Butanone	6.84	43	99542	43.034	ug/l	99
29) Tetrahydrofuran	7.21	42	172496	115.310	ug/l	99
44) Trichloroethene	8.84	130	3722	0.453	ug/l	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022119\
Data File : VN053886.D
Acq On : 21 Feb 2019 12:29
Operator : JC/SP
Sample : K1540-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
M2-T2-OZ(PRE)

Quant Time: Feb 22 01:12:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022019W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 21 03:36:07 2019
Response via : Initial Calibration

