

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022119\
 Data File : VN053893.D
 Acq On : 21 Feb 2019 15:37
 Operator : JC/SP
 Sample : K1344-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NB-08-022019-A

Quant Time: Feb 22 02:14:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 21 03:36:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	619303	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1053178	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	988599	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	386972	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	334585	52.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.72%	
35) Dibromofluoromethane	7.59	113	334942	48.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.92%	
50) Toluene-d8	10.09	98	1281644	51.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.68%	
62) 4-Bromofluorobenzene	12.40	95	420026	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
Target Compounds						
16) Acetone	3.82	43	23342	13.410	ug/l	94
20) Methylene Chloride	4.56	84	50121	7.691	ug/l	99
43) Isopropyl Acetate	8.18	43	27364	2.657	ug/l #	92
95) Naphthalene	15.13	128	78599	9.259	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022119\
Data File : VN053893.D
Acq On : 21 Feb 2019 15:37
Operator : JC/SP
Sample : K1344-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NB-08-022019-A

Quant Time: Feb 22 02:14:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022019W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 21 03:36:07 2019
Response via : Initial Calibration

