

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022119\
 Data File : VN053896.D
 Acq On : 21 Feb 2019 16:57
 Operator : JC/SP
 Sample : K1538-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 M-11-A-S

Quant Time: Feb 22 02:24:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 21 03:36:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	616655	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1045367	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1023233	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	390026	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	330914	52.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.02%	
35) Dibromofluoromethane	7.59	113	333037	48.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.08%	
50) Toluene-d8	10.09	98	1273483	51.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.78%	
62) 4-Bromofluorobenzene	12.40	95	424057	51.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.68%	
Target Compounds						
16) Acetone	3.82	43	26487	15.282	ug/l	97
20) Methylene Chloride	4.56	84	742867	114.484	ug/l	99
25) 2-Butanone	6.84	43	4874	2.179	ug/l	90
43) Isopropyl Acetate	8.18	43	31594	3.091	ug/l #	89
95) Naphthalene	15.13	128	3582	4.822	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022119\
Data File : VN053896.D
Acq On : 21 Feb 2019 16:57
Operator : JC/SP
Sample : K1538-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
M-11-A-S

Quant Time: Feb 22 02:24:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022019W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 21 03:36:07 2019
Response via : Initial Calibration

