

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022119\
 Data File : VN053899.D
 Acq On : 21 Feb 2019 18:17
 Operator : JC/SP
 Sample : K1349-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-02-021919-B

Quant Time: Feb 22 02:38:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 21 03:36:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	674012	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1144589	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1073157	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	417286	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	362167	52.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.16%	
35) Dibromofluoromethane	7.59	113	365453	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
50) Toluene-d8	10.09	98	1376932	50.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.50%	
62) 4-Bromofluorobenzene	12.40	95	448052	50.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.04%	
Target Compounds						
16) Acetone	3.82	43	47769	25.216	ug/l	96
20) Methylene Chloride	4.56	84	192824	27.187	ug/l	98
43) Isopropyl Acetate	8.18	43	37103	3.315	ug/l #	85
52) Toluene	10.16	92	24429	1.329	ug/l	96
95) Naphthalene	15.13	128	18149	5.607	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022119\
Data File : VN053899.D
Acq On : 21 Feb 2019 18:17
Operator : JC/SP
Sample : K1349-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-02-021919-B

Quant Time: Feb 22 02:38:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022019W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 21 03:36:07 2019
Response via : Initial Calibration

