

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022119\
 Data File : VN053901.D
 Acq On : 21 Feb 2019 19:11
 Operator : JC/SP
 Sample : K1546-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RT-3422

Manual Integrations
 APPROVED

MMDadoda
 2/22/2019 10:30:57 AM

Quant Time: Feb 22 02:48:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 21 03:36:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	608869	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1034645	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	973615	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	382581	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	324103	52.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.18%	
35) Dibromofluoromethane	7.59	113	327739	48.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.54%	
50) Toluene-d8	10.09	98	1255782	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
62) 4-Bromofluorobenzene	12.40	95	419086	51.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.52%	
Target Compounds						
16) Acetone	3.82	43	16588	9.693	ug/l	100
20) Methylene Chloride	4.56	84	58473	9.127	ug/l	95
43) Isopropyl Acetate	8.18	43	21464m	2.122	ug/l	
95) Naphthalene	15.14	128	5184	4.922	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022119\
Data File : VN053901.D
Acq On : 21 Feb 2019 19:11
Operator : JC/SP
Sample : K1546-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RT-3422

Manual Integrations
APPROVED

MMDadoda
2/22/2019 10:30:57 AM

Quant Time: Feb 22 02:48:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022019W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 21 03:36:07 2019
Response via : Initial Calibration

