

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022120\
 Data File : VN060187.D
 Acq On : 21 Feb 2020 14:00
 Operator : JC/MD
 Sample : L1645-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Feb 21 23:05:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	177910	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	281608	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	262502	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	113067	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	105477	48.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.44%	
35) Dibromofluoromethane	7.56	113	84372	48.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.48%	
50) Toluene-d8	10.08	98	337250	49.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.94%	
62) 4-Bromofluorobenzene	12.40	95	118987	47.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.00%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022120\
Data File : VN060187.D
Acq On : 21 Feb 2020 14:00
Operator : JC/MD
Sample : L1645-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Feb 21 23:05:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

