

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022123\
 Data File : VN076746.D
 Acq On : 21 Feb 2023 11:00
 Operator : JC\MD
 Sample : VN0221WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0221WBL01

Quant Time: Feb 22 04:10:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	274936	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	500008	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	434825	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	165250	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	204779	53.939	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.880%
35) Dibromofluoromethane	8.172	113	154087	47.057	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.120%
50) Toluene-d8	10.572	98	602349	49.257	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.520%
62) 4-Bromofluorobenzene	12.854	95	174045	42.972	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022123\
Data File : VN076746.D
Acq On : 21 Feb 2023 11:00
Operator : JC\MD
Sample : VN0221WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0221WBL01

Quant Time: Feb 22 04:10:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

