

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022123\
 Data File : VN076757.D
 Acq On : 21 Feb 2023 18:01
 Operator : JC\MD
 Sample : 01641-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B-26-GW01

Quant Time: Feb 22 04:13:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	249287	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	437750	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	392280	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	150756	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	177194	51.476	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.960%
35) Dibromofluoromethane	8.172	113	137023	47.797	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.600%
50) Toluene-d8	10.572	98	548128	51.198	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.400%
62) 4-Bromofluorobenzene	12.854	95	156961	44.266	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.540%
Target Compounds						
						Qvalue
16) Acetone	4.449	43	16676	10.845	ug/l #	77
25) 2-Butanone	7.490	43	3870	1.706	ug/l	94
52) Toluene	10.625	92	2697	0.351	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022123\
Data File : VN076757.D
Acq On : 21 Feb 2023 18:01
Operator : JC\MD
Sample : 01641-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B-26-GW01

Quant Time: Feb 22 04:13:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

