

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022124\
 Data File : VN081109.D
 Acq On : 21 Feb 2024 10:39
 Operator : JC\MD
 Sample : VN0221WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0221WBL01

Quant Time: Feb 22 04:42:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	300144	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	527062	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	452721	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	191388	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	162149	44.075	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.160%
35) Dibromofluoromethane	8.171	113	138646	46.520	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.040%
50) Toluene-d8	10.565	98	454948	42.978	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	85.960%#
62) 4-Bromofluorobenzene	12.847	95	163242	41.041	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.080%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022124\
Data File : VN081109.D
Acq On : 21 Feb 2024 10:39
Operator : JC\MD
Sample : VN0221WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0221WBL01

Quant Time: Feb 22 04:42:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

