

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022124\
 Data File : VN081115.D
 Acq On : 21 Feb 2024 13:11
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Feb 22 04:45:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	315022	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	615188	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	523168	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	212341	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	201477	52.179	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.360%
35) Dibromofluoromethane	8.171	113	166323	47.812	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.620%
50) Toluene-d8	10.565	98	531311	43.002	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	86.000%
62) 4-Bromofluorobenzene	12.847	95	191819	41.317	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.640%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022124\
Data File : VN081115.D
Acq On : 21 Feb 2024 13:11
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Feb 22 04:45:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

