

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022124\
 Data File : VN081117.D
 Acq On : 21 Feb 2024 13:59
 Operator : JC\MD
 Sample : P1540-05 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1327

Manual Integrations
 APPROVED

Reviewed By : John Carlone 02/22/2024
 Supervised By : Semsettin Yesilyurt 02/22/2024

Quant Time: Feb 22 04:46:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.223	168	323506	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	610798	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	517912	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	214767	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	204488	51.570	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.140%			
35) Dibromofluoromethane	8.171	113	172273	49.878	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.760%			
50) Toluene-d8	10.565	98	580994	47.361	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 94.720%			
62) 4-Bromofluorobenzene	12.853	95	209994	45.557	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 91.120%			
Target Compounds						
16) Acetone	4.424	43	3690m	2.365	ug/l	
40) Benzene	8.606	78	6702	0.391	ug/l #	83
52) Toluene	10.635	92	3569	0.329	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022124\
Data File : VN081117.D
Acq On : 21 Feb 2024 13:59
Operator : JC\MD
Sample : P1540-05 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1327

Quant Time: Feb 22 04:46:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/22/2024
Supervised By :Semsettin Yesilyurt 02/22/2024

