

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022124\
 Data File : VN081120.D
 Acq On : 21 Feb 2024 15:10
 Operator : JC\MD
 Sample : P1546-08 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1331

Quant Time: Feb 22 04:47:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	332016	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	613008	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	531754	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	219839	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	185067	45.476	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	90.960%		
35) Dibromofluoromethane	8.171	113	156888	45.260	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.520%		
50) Toluene-d8	10.571	98	525728	42.702	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	85.400%#		
62) 4-Bromofluorobenzene	12.847	95	188073	40.654	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	81.300%#		
Target Compounds						
52) Toluene	10.629	92	3042	0.280	ug/l	95
68) m/p-Xylenes	12.070	106	3648	0.462	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022124\
Data File : VN081120.D
Acq On : 21 Feb 2024 15:10
Operator : JC\MD
Sample : P1546-08 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1331

Quant Time: Feb 22 04:47:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

