

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022125\
 Data File : VN085804.D
 Acq On : 21 Feb 2025 15:42
 Operator : JC\MD
 Sample : VN0221WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0221WBL01

Quant Time: Feb 22 02:35:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	239124	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	440255	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	393842	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	150905	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	186621	60.197	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	120.400%
35) Dibromofluoromethane	8.159	113	164740	57.185	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	114.380%
50) Toluene-d8	10.565	98	530599	50.623	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.240%
62) 4-Bromofluorobenzene	12.847	95	152825	44.254	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	88.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022125\
Data File : VN085804.D
Acq On : 21 Feb 2025 15:42
Operator : JC\MD
Sample : VN0221WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0221WBL01

Quant Time: Feb 22 02:35:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

