

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022125\
 Data File : VN085819.D
 Acq On : 21 Feb 2025 21:50
 Operator : JC\MD
 Sample : Q1397-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP8

Quant Time: Feb 22 02:43:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	245873	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	488846	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	463054	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	164566	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	194272	60.945	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 121.900%	
35) Dibromofluoromethane	8.165	113	166064	51.915	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.820%	
50) Toluene-d8	10.565	98	631804	54.287	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.580%	
62) 4-Bromofluorobenzene	12.847	95	183039	47.735	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 95.460%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	41830	41.472	ug/l	96
18) Methyl Acetate	5.030	43	4025	1.160	ug/l #	88
20) Methylene Chloride	5.283	84	4513	1.391	ug/l #	68
37) Ethyl Acetate	7.553	43	4331	1.163	ug/l #	80
43) Isopropyl Acetate	8.682	43	210792	32.043	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022125\
Data File : VN085819.D
Acq On : 21 Feb 2025 21:50
Operator : JC\MD
Sample : Q1397-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP8

Quant Time: Feb 22 02:43:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

