

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022219\
 Data File : VN053907.D
 Acq On : 22 Feb 2019 11:18
 Operator : JC/SP
 Sample : PB117286TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB117286TB

Quant Time: Feb 22 22:32:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 21 03:36:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	614588	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1094537	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1010609	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	376416	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	341036	54.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.60%	
35) Dibromofluoromethane	7.59	113	334245	46.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.06%	
50) Toluene-d8	10.09	98	1345799	51.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.74%	
62) 4-Bromofluorobenzene	12.40	95	420929	49.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.28%	
Target Compounds						
16) Acetone	3.82	43	18472	10.694	ug/l	91
20) Methylene Chloride	4.56	84	18474	2.857	ug/l	96
43) Isopropyl Acetate	8.17	43	32908	3.075	ug/l #	89
95) Naphthalene	15.14	128	5072	4.921	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022219\
Data File : VN053907.D
Acq On : 22 Feb 2019 11:18
Operator : JC/SP
Sample : PB117286TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB117286TB

Quant Time: Feb 22 22:32:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022019W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 21 03:36:07 2019
Response via : Initial Calibration

