

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022219\
 Data File : VN053913.D
 Acq On : 22 Feb 2019 13:58
 Operator : JC/SP
 Sample : K1546-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RT-1495

Quant Time: Feb 22 22:53:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 21 03:36:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	595166	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1075094	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1072846	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	347330	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	324203	53.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.60%	
35) Dibromofluoromethane	7.59	113	325714	46.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.32%	
50) Toluene-d8	10.09	98	1346755	52.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.70%	
62) 4-Bromofluorobenzene	12.40	95	393601	46.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.56%	
Target Compounds						
16) Acetone	3.82	43	19808	11.841	ug/l	99
20) Methylene Chloride	4.55	84	2792641	445.915	ug/l	100
43) Isopropyl Acetate	8.18	43	27659	2.631	ug/l #	90
95) Naphthalene	15.13	128	16093	5.672	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022219\
Data File : VN053913.D
Acq On : 22 Feb 2019 13:58
Operator : JC/SP
Sample : K1546-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RT-1495

Quant Time: Feb 22 22:53:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022019W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 21 03:36:07 2019
Response via : Initial Calibration

