

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022221\
 Data File : VN065955.D
 Acq On : 22 Feb 2021 13:22
 Operator : JC/MD
 Sample : M1450-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-TT-VPB-RW6-EB-20210216

Quant Time: Feb 22 15:12:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.631	168	172011	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	283538	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	259340	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	106193	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	109461	45.51	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.02%
35) Dibromofluoromethane	7.553	113	83722	47.71	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.42%
50) Toluene-d8	10.058	98	345048	47.90	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.80%
62) 4-Bromofluorobenzene	12.373	95	123127	45.53	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.06%
Target Compounds						
16) Acetone	3.798	43	9578	11.14	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022221\
Data File : VN065955.D
Acq On : 22 Feb 2021 13:22
Operator : JC/MD
Sample : M1450-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-TT-VPB-RW6-EB-20210216

Quant Time: Feb 22 15:12:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

