

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022221\
 Data File : VN065962.D
 Acq On : 22 Feb 2021 16:08
 Operator : JC/MD
 Sample : M1431-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-02-021621

Quant Time: Feb 23 00:09:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	152114	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	250422	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	228350	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	96144	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.990	65	98635	46.37	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.74%
35) Dibromofluoromethane	7.553	113	73902	47.68	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.36%
50) Toluene-d8	10.058	98	305857	48.08	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.16%
62) 4-Bromofluorobenzene	12.373	95	111683	46.76	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.52%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022221\
Data File : VN065962.D
Acq On : 22 Feb 2021 16:08
Operator : JC/MD
Sample : M1431-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-02-021621

Quant Time: Feb 23 00:09:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

