

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022221\
 Data File : VN065971.D
 Acq On : 22 Feb 2021 19:42
 Operator : JC/MD
 Sample : M1452-24
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-6

Quant Time: Feb 23 00:13:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	153759	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	252961	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	233584	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	98038	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.990	65	99888	46.46	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.92%
35) Dibromofluoromethane	7.553	113	76339	48.76	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.52%
50) Toluene-d8	10.058	98	309456	48.15	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.30%
62) 4-Bromofluorobenzene	12.373	95	113967	47.24	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.48%
Target Compounds						
16) Acetone	3.791	43	25751	36.76	ug/l	98
20) Methylene Chloride	4.518	84	6682	3.35	ug/l	94
43) Isopropyl Acetate	8.126	43	8275	2.31	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022221\
Data File : VN065971.D
Acq On : 22 Feb 2021 19:42
Operator : JC/MD
Sample : M1452-24
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-6

Quant Time: Feb 23 00:13:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

