

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022221\
 Data File : VN065972.D
 Acq On : 22 Feb 2021 20:05
 Operator : JC/MD
 Sample : M1452-28
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-7

Quant Time: Feb 23 00:13:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	168954	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	281851	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	258516	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	110515	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	110230	46.65	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.30%
35) Dibromofluoromethane	7.553	113	83361	47.79	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.58%
50) Toluene-d8	10.058	98	342916	47.89	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.78%
62) 4-Bromofluorobenzene	12.373	95	127677	47.50	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.00%
Target Compounds						
					Qvalue	
16) Acetone	3.791	43	16923	21.33	ug/l	97
20) Methylene Chloride	4.521	84	6642	3.03	ug/l	92
43) Isopropyl Acetate	8.129	43	9712	2.43	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022221\
Data File : VN065972.D
Acq On : 22 Feb 2021 20:05
Operator : JC/MD
Sample : M1452-28
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-7

Quant Time: Feb 23 00:13:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

