

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022222\
 Data File : VN071021.D
 Acq On : 22 Feb 2022 11:58
 Operator : JC\MD
 Sample : VN0222WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0222WBL01

Quant Time: Feb 23 00:57:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	887706	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1406407	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1229356	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	408493	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	525285	51.801	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.600%
35) Dibromofluoromethane	8.016	113	408787	48.867	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.740%
50) Toluene-d8	10.439	98	1676052	49.891	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.780%
62) 4-Bromofluorobenzene	12.727	95	548240	45.656	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022222\
Data File : VN071021.D
Acq On : 22 Feb 2022 11:58
Operator : JC\MD
Sample : VN0222WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0222WBL01

Quant Time: Feb 23 00:57:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

