

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022222\
 Data File : VN071024.D
 Acq On : 22 Feb 2022 13:34
 Operator : JC\MD
 Sample : N1629-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-TT-MW161S1-GW-220-222

Quant Time: Feb 23 00:58:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	818147	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1307475	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1131347	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	389627	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	492860	52.736	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	105.480%
35) Dibromofluoromethane	8.022	113	384347	49.422	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.840%
50) Toluene-d8	10.439	98	1546094	49.505	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.000%
62) 4-Bromofluorobenzene	12.727	95	512186	45.881	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.760%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	14648	2.846	ug/l	95
24) 1,1-Dichloroethane	6.387	63	36535	2.225	ug/l	97
95) Naphthalene	15.504	128	20659	6.057	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022222\
Data File : VN071024.D
Acq On : 22 Feb 2022 13:34
Operator : JC\MD
Sample : N1629-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-TT-MW161S1-GW-220-222

Quant Time: Feb 23 00:58:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

