

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022223\
 Data File : VN076779.D
 Acq On : 22 Feb 2023 18:28
 Operator : JC\MD
 Sample : 01657-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-2023-2-15

Quant Time: Feb 22 23:59:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.237	168	271158	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.101	114	487884	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	422116	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	164296	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	199748	53.347	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.700%
35) Dibromofluoromethane	8.178	113	153551	48.059	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.120%
50) Toluene-d8	10.572	98	595174	49.880	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.760%
62) 4-Bromofluorobenzene	12.854	95	176042	44.545	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022223\
Data File : VN076779.D
Acq On : 22 Feb 2023 18:28
Operator : JC\MD
Sample : 01657-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-2023-2-15

Quant Time: Feb 22 23:59:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

