

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022224\
 Data File : VN081137.D
 Acq On : 22 Feb 2024 15:59
 Operator : JC\MD
 Sample : PB159161TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB159161TB

Quant Time: Feb 23 04:27:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	339829	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	648529	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	583320	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	234873	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	191760	46.037	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.080%
35) Dibromofluoromethane	8.165	113	162811	44.396	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.800%
50) Toluene-d8	10.571	98	538026	41.307	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	82.620%#
62) 4-Bromofluorobenzene	12.853	95	197181	40.289	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	80.580%#
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	65093	39.716	ug/l	94
20) Methylene Chloride	5.283	84	49639	10.982	ug/l #	86
37) Ethyl Acetate	7.559	43	7840	1.383	ug/l	95
43) Isopropyl Acetate	8.688	43	500467	50.466	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022224\
Data File : VN081137.D
Acq On : 22 Feb 2024 15:59
Operator : JC\MD
Sample : PB159161TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB159161TB

Quant Time: Feb 23 04:27:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

