

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022224\
 Data File : VN081141.D
 Acq On : 22 Feb 2024 17:35
 Operator : JC\MD
 Sample : P1539-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20241105-AMENDED-COMPOSITE-35N-N-4-04

Quant Time: Feb 23 04:29:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	308604	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	596150	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	513068	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	207607	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	191071	50.513	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.020%
35) Dibromofluoromethane	8.177	113	73	0.022	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.040%#
50) Toluene-d8	10.571	98	508820	42.497	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	85.000%#
62) 4-Bromofluorobenzene	12.847	95	188947	41.998	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.000%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	80179	53.870	ug/l	95
20) Methylene Chloride	5.283	84	11951	2.215	ug/l #	84
25) 2-Butanone	7.488	43	18032	7.499	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022224\
Data File : VN081141.D
Acq On : 22 Feb 2024 17:35
Operator : JC\MD
Sample : P1539-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20241105-AMENDED-COMPOSITE-35N-N-4-04

Quant Time: Feb 23 04:29:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

