

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022224\
 Data File : VN081144.D
 Acq On : 22 Feb 2024 18:47
 Operator : JC\MD
 Sample : PB159186TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB159186TB

Manual Integrations
 APPROVED

Reviewed By : John Carlone 02/23/2024
 Supervised By : Mahesh Dadoda 02/23/2024

Quant Time: Feb 23 04:30:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	293164	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	572477	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	487796	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	193426	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	189696	52.791	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.580%			
35) Dibromofluoromethane	8.171	113	151643	46.844	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 93.680%			
50) Toluene-d8	10.571	98	479322	41.689	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 83.380%#			
62) 4-Bromofluorobenzene	12.853	95	176910	40.949	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 81.900%#			
Target Compounds						
16) Acetone	4.447	43	9260m	6.549	ug/l	
20) Methylene Chloride	5.283	84	8450	1.407	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022224\
Data File : VN081144.D
Acq On : 22 Feb 2024 18:47
Operator : JC\MD
Sample : PB159186TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB159186TB

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/23/2024
Supervised By :Mahesh Dadoda 02/23/2024

Quant Time: Feb 23 04:30:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

