

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022224\
 Data File : VN081145.D
 Acq On : 22 Feb 2024 19:11
 Operator : JC\MD
 Sample : P1549-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HT3727

Quant Time: Feb 23 04:31:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	265183	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	525090	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	462656	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	180850	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	193089	59.405	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	118.800%	
35) Dibromofluoromethane	8.171	113	160439	54.034	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	108.060%	
50) Toluene-d8	10.571	98	513800	48.720	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	97.440%	
62) 4-Bromofluorobenzene	12.853	95	189837	47.907	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	95.820%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	85960	67.211	ug/l	100
20) Methylene Chloride	5.283	84	40115	11.407	ug/l #	83
37) Ethyl Acetate	7.565	43	7590	1.653	ug/l #	92
43) Isopropyl Acetate	8.694	43	237149	29.535	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022224\
Data File : VN081145.D
Acq On : 22 Feb 2024 19:11
Operator : JC\MD
Sample : P1549-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HT3727

Quant Time: Feb 23 04:31:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

