

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022321\
 Data File : VN065977.D
 Acq On : 23 Feb 2021 11:28
 Operator : JC/MD
 Sample : VN0223MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0223MBL01

Quant Time: Feb 24 00:31:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	151129	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	244755	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	219795	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	86408	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	94874	44.89	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	89.78%
35) Dibromofluoromethane	7.550	113	73563	48.56	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.12%
50) Toluene-d8	10.055	98	302235	48.61	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.22%
62) 4-Bromofluorobenzene	12.373	95	103848	44.49	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.98%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022321\
Data File : VN065977.D
Acq On : 23 Feb 2021 11:28
Operator : JC/MD
Sample : VN0223MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0223MBL01

Quant Time: Feb 24 00:31:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

