

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022321\
 Data File : VN065981.D
 Acq On : 23 Feb 2021 13:19
 Operator : JC/MD
 Sample : PB134775TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134775TB

Quant Time: Feb 24 00:33:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	144525	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	238009	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	217743	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	89374	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.990	65	93225	46.13	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.26%
35) Dibromofluoromethane	7.553	113	71635	48.63	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.26%
50) Toluene-d8	10.058	98	293417	48.53	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.06%
62) 4-Bromofluorobenzene	12.373	95	104450	46.02	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.04%
Target Compounds						
16) Acetone	3.791	43	25228	38.38	ug/l	99
20) Methylene Chloride	4.521	84	8644	4.61	ug/l	95
43) Isopropyl Acetate	8.129	43	10181	3.01	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022321\
Data File : VN065981.D
Acq On : 23 Feb 2021 13:19
Operator : JC/MD
Sample : PB134775TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134775TB

Quant Time: Feb 24 00:33:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

