

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN022321\  
 Data File : VN065982.D  
 Acq On : 23 Feb 2021 13:41  
 Operator : JC/MD  
 Sample : PB134775ZHE#01  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 PB134775ZHE#01

Quant Time: Feb 24 00:34:09 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N021721W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Feb 17 15:48:13 2021  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response   | Conc   | Units  | Dev(Min) |
|-----------------------------|----------------|------|------------|--------|--------|----------|
| -----                       |                |      |            |        |        |          |
| Internal Standards          |                |      |            |        |        |          |
| 1) Pentafluorobenzene       | 7.630          | 168  | 146417     | 50.00  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.553          | 114  | 240747     | 50.00  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 11.380         | 117  | 217783     | 50.00  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.315         | 152  | 90794      | 50.00  | ug/l   | 0.00     |
|                             |                |      |            |        |        |          |
| System Monitoring Compounds |                |      |            |        |        |          |
| 33) 1,2-Dichloroethane-d4   | 7.991          | 65   | 93274      | 45.55  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 61 - 141 |      | Recovery = | 91.10% |        |          |
| 35) Dibromofluoromethane    | 7.553          | 113  | 72231      | 48.48  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 69 - 133 |      | Recovery = | 96.96% |        |          |
| 50) Toluene-d8              | 10.058         | 98   | 292420     | 47.81  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 65 - 126 |      | Recovery = | 95.62% |        |          |
| 62) 4-Bromofluorobenzene    | 12.373         | 95   | 104447     | 45.49  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 58 - 135 |      | Recovery = | 90.98% |        |          |
|                             |                |      |            |        |        |          |
| Target Compounds            |                |      |            |        |        |          |
| 16) Acetone                 | 3.795          | 43   | 25966      | 39.02  | ug/l   | 95       |
| 20) Methylene Chloride      | 4.521          | 84   | 8193       | 4.32   | ug/l   | 90       |
| 43) Isopropyl Acetate       | 8.129          | 43   | 10310      | 3.02   | ug/l # | 91       |
| -----                       |                |      |            |        |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN022321\  
Data File : VN065982.D  
Acq On : 23 Feb 2021 13:41  
Operator : JC/MD  
Sample : PB134775ZHE#01  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
PB134775ZHE#01

Quant Time: Feb 24 00:34:09 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N021721W.M  
Quant Title : SW846 8260  
QLast Update : Wed Feb 17 15:48:13 2021  
Response via : Initial Calibration

