

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022321\
 Data File : VN065984.D
 Acq On : 23 Feb 2021 14:27
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Feb 24 00:34:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	139590	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	229261	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	209118	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	84699	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.990	65	89111	45.65	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.30%
35) Dibromofluoromethane	7.553	113	68263	48.11	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.22%
50) Toluene-d8	10.058	98	275814	47.36	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.72%
62) 4-Bromofluorobenzene	12.373	95	100223	45.84	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.68%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022321\
Data File : VN065984.D
Acq On : 23 Feb 2021 14:27
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Feb 24 00:34:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

