

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022321\
 Data File : VN065988.D
 Acq On : 23 Feb 2021 16:02
 Operator : JC/MD
 Sample : M1470-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CBH-139

Quant Time: Feb 24 00:36:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.631	168	138188	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	230399	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	207461	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	85285	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	89996	46.57	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.14%
35) Dibromofluoromethane	7.553	113	68916	48.33	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.66%
50) Toluene-d8	10.058	98	279494	47.75	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.50%
62) 4-Bromofluorobenzene	12.373	95	101977	46.41	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.82%
Target Compounds						
					Qvalue	
30) Chloroform	7.335	83	33243	10.99	ug/l	96
47) Bromodichloromethane	9.370	83	2122	0.96	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022321\
Data File : VN065988.D
Acq On : 23 Feb 2021 16:02
Operator : JC/MD
Sample : M1470-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CBH-139

Quant Time: Feb 24 00:36:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

