

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022321\
 Data File : VN065989.D
 Acq On : 23 Feb 2021 16:26
 Operator : JC/MD
 Sample : M1470-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CBH-140

Quant Time: Feb 24 00:36:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	136655	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	225194	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	205024	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	83585	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	91657	47.96	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.92%
35) Dibromofluoromethane	7.553	113	68470	49.13	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.26%
50) Toluene-d8	10.058	98	273985	47.89	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.78%
62) 4-Bromofluorobenzene	12.376	95	100921	46.99	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.98%
Target Compounds						
					Qvalue	
30) Chloroform	7.335	83	60091	20.09	ug/l	95
47) Bromodichloromethane	9.370	83	4522	2.09	ug/l #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022321\
Data File : VN065989.D
Acq On : 23 Feb 2021 16:26
Operator : JC/MD
Sample : M1470-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CBH-140

Quant Time: Feb 24 00:36:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

