

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022322\
 Data File : VN071033.D
 Acq On : 23 Feb 2022 13:17
 Operator : JC\MD
 Sample : VN0223WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0223WBL01

Quant Time: Feb 24 00:39:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	874625	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1391151	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1229966	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	404827	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	521192	52.166	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.340%
35) Dibromofluoromethane	8.022	113	405452	49.000	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.000%
50) Toluene-d8	10.439	98	1667328	50.175	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.360%
62) 4-Bromofluorobenzene	12.727	95	550517	46.348	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022322\
Data File : VN071033.D
Acq On : 23 Feb 2022 13:17
Operator : JC\MD
Sample : VN0223WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0223WBL01

Quant Time: Feb 24 00:39:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

