

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022322\
 Data File : VN071040.D
 Acq On : 23 Feb 2022 18:08
 Operator : JC\MD
 Sample : N1642-03 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31455

Quant Time: Feb 24 00:41:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	806897	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1282094	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1135824	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	401520	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	487695	52.911	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 105.820%			
35) Dibromofluoromethane	8.022	113	378842	49.678	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 99.360%			
50) Toluene-d8	10.439	98	1546549	50.500	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 101.000%			
62) 4-Bromofluorobenzene	12.727	95	526617	48.107	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 96.220%			
Target Compounds					Qvalue	
16) Acetone	4.292	43	20829	4.103	ug/l	93
51) 4-Methyl-2-Pentanone	10.327	43	5691	0.443	ug/l	96
52) Toluene	10.504	92	88854	3.826	ug/l	96
68) m/p-Xylenes	11.945	106	12532	0.757	ug/l	100
69) o-Xylene	12.280	106	4765	0.295	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022322\
Data File : VN071040.D
Acq On : 23 Feb 2022 18:08
Operator : JC\MD
Sample : N1642-03 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31455

Quant Time: Feb 24 00:41:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

