

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022322\
 Data File : VN071041.D
 Acq On : 23 Feb 2022 18:33
 Operator : JC\MD
 Sample : N1642-04 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31456

Quant Time: Feb 24 00:41:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	859723	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1350494	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1208308	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	421935	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	510134	51.944	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.880%
35) Dibromofluoromethane	8.022	113	398529	49.613	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.220%
50) Toluene-d8	10.439	98	1646284	51.034	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.060%
62) 4-Bromofluorobenzene	12.727	95	554433	48.083	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022322\
Data File : VN071041.D
Acq On : 23 Feb 2022 18:33
Operator : JC\MD
Sample : N1642-04 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31456

Quant Time: Feb 24 00:41:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

