

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022322\
 Data File : VN071043.D
 Acq On : 23 Feb 2022 19:21
 Operator : JC\MD
 Sample : N1642-06 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31458

Quant Time: Feb 24 00:41:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	876010	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1373782	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1230546	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	418554	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	523434	52.308	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.620%
35) Dibromofluoromethane	8.016	113	402636	49.275	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.540%
50) Toluene-d8	10.439	98	1676041	51.075	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.160%
62) 4-Bromofluorobenzene	12.727	95	558262	47.595	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022322\
Data File : VN071043.D
Acq On : 23 Feb 2022 19:21
Operator : JC\MD
Sample : N1642-06 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31458

Quant Time: Feb 24 00:41:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

