

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022322\
 Data File : VN071044.D
 Acq On : 23 Feb 2022 19:45
 Operator : JC\MD
 Sample : N1642-08 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31409

Quant Time: Feb 24 00:42:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	822969	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1295363	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1150754	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	379345	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.439	65	495009	52.655	ug/l	0.01
Spiked Amount	50.000	Range	61 - 141	Recovery	= 105.320%	
35) Dibromofluoromethane	8.022	113	381632	49.532	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 99.060%	
50) Toluene-d8	10.439	98	1582436	51.142	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.280%	
62) 4-Bromofluorobenzene	12.727	95	516688	46.717	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 93.440%	
Target Compounds						
						Qvalue
16) Acetone	4.293	43	61399	11.859	ug/l	93
52) Toluene	10.504	92	111196	4.739	ug/l	98
68) m/p-Xylenes	11.945	106	8321	0.496	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022322\
Data File : VN071044.D
Acq On : 23 Feb 2022 19:45
Operator : JC\MD
Sample : N1642-08 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31409

Quant Time: Feb 24 00:42:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

